

ACCOMPANYING LABELING: Folder entitled "Protect Your Health With Relax-O-Thermassage."

RESULTS OF INVESTIGATION: The device was a cushioned box containing a motor capable of providing vibrating motions.

LIBELED: 5-12-58, E. Dist. Pa.

CHARGE: 502(a)—when shipped, the labeling contained false and misleading representations that the device was an adequate and effective treatment for overcoming nervous tension, muscular stiffness, aching joints and muscles; increasing blood circulation; aiding the body to make use of its own protective and self healing powers; to rid the body of wastes, poisons, and foreign matter; to stimulate organs of purification and elimination, thereby freeing nerves, blood vessels, and lymph of obstructions; equalizing circulation throughout all parts of the body; restoring nerve energy through proper blood supply to diseased organs and tissues; overcoming insomnia, arthritic joints, neuritis, and lumbago; and providing effective figure control.

DISPOSITION: Relax-O-Therapy Equipment Co., Inc., Dearborn, Mich., appeared as claimant and denied that the device was misbranded. On 9-15-59, the Government filed written interrogatories. The claimant failed to answer the interrogatories and thereafter, on 3-16-60, a default decree of condemnation was entered and the article was destroyed.

6071. Electro-Galvanic Apparatus. (F.D.C. No. 41893. S. No. 6-580 P.)

QUANTITY: 50 devices and one carton of device accessories at Waltham, Mass.

SHIPPED: The devices were shipped on 8-20-56, from Hamburg, Germany, by Actien-Gesellschaft, and the accessories were shipped on 11-5-57, from Hamburg, Germany, by Caesar Kock Nachf.

LABEL IN PART: (Device) "Wolmuth-Zentrale * * * Original-Apparat No. 125027" or "Elga 10093 Kosmetische Elga-Zentrale"; (ctn. of accessories) "Hammond & Montgomery * * * Boston/USA."

ACCOMPANYING LABELING: Book entitled "Die Elektrogalvanische Heilkunde."

RESULTS OF INVESTIGATION: The article consisted of dry cells, a meter, and a variety of electrodes for applying galvanic current to all parts of the body.

LIBELED: 6-25-58, Dist. Mass.

CHARGE: 502(a)—when shipped and while held for sale, the labeling which accompanied the device contained false and misleading representations that the device was an adequate and effective treatment for conditions of the eyes, urinary tract, chest, feet, elbow, gallbladder, palate, joints, throat, heart, larynx, liver, stomach, rectum, nasal passages, ears, kidney, prostate, spinal cord, and vertebral column; asthma, bronchitis, hemorrhoids, varicose veins, goiter, paralysis, and pain.

DISPOSITION: On 5-5-59, George W. A. Hammond, Duxbury, Mass., appeared as claimant. Thereafter, on 9-9-59, the Government filed written interrogatories. After the claimant failed to answer the interrogatories, a default decree was entered on 2-1-60, and the device was destroyed.

6072. Trim-Tone vibrating device. (F.D.C. No. 41931. S. Nos. 31-552/5 P.)

QUANTITY: 150 Model SX devices, 12 Model T devices, 50 Model A devices, and 12 Model D devices, at Jamaica, N.Y., and 36 Model SX devices, 6 Model T devices, 6 Model A devices, and 6 Model D devices at Flushing, N.Y., in possession of B. Gertz, Inc.